

NO. CPRGHK/STORE/PLAN QTN/
Office of the Dean,
Chhatrapati Pramila Rajee General
Hospital, Kolhapur.
Dt. 30/01/2020

1182
/19-20

To,
President,
Website Publishing Committee and
Professor And H.O.D., Dept. of P.S.M.,
C.P.R. General Hospital, Kolhapur.

Sub : Quotation for Machinery and Equipments.
Ref. : As per Hon. Deans approved note Dated : 30/01/2020

Sir,

Please quote your competitive rates in Indian Rupees only for the following item. This quotation procedure is subject to permission / administrative approval from Government of Maharashtra.

Specification No.	Name of Machinery (User Department)	Specification :	Rate Per Unit
PL - 44	Fiber Optic Cable For Light Source	Fiber optics cable 5meter length flexible moment superior nano optics Natural day light out source comfortable with all light source	
PL - 45	Pulse Oxymeter	<ol style="list-style-type: none"> 1. It should provide Spo2 and Pulse rate measurements of any patient ranging from Adult and Neonates. 2. Option to display plethysmograph or Values in extra Big Digits. 3. Performance in good. 4. Audible Pitch change with change in % SpO2 5. Non-volatile memory for storing Graphical & Tabular trends with alarm status for min.24 Hours 6. Serial interface for connecting printer / PC for printing Spo2 and PR values every 1 Min, 5min, 15min, 30min 7. 7. Compatible to connect CMS software. 8. Long battery operation using power saving mode 	
PL - 46	Psychanalytic test	<p>1. Thematic Apperception Test- Indian Adaptation Uma Choudhary (TAT) Specification It is a method of revealing to the trained interpreter some of the dominant drives, emotions, sentiments, complexes and conflict of a personality. It is found useful in comprehensive study of personality,behaviour disorders,psychosomatic illness etc.</p> <ol style="list-style-type: none"> 1. It has 31 cards- 2. 10 for male, 3. 10 for female and 4. 10common and one blank card, 5. one has to write story after seeing card. <p>2. Children apperception test (C.A.T.) Specification This projective technique presents situations of special concern to children. It consists of 10 animal pictures in a social context involving the child in conflict, identities, roles, family structures, and interpersonal interaction. The supplement (C.A.T.-S) presents children with common family situations (i.e., prolonged illness, physical disability, mother's pregnancy, separation of parents). Children who do</p>	

		<p>not produce stories readily can manipulate these forms as a play technique. The C.A.T.-H consists of human figures and situations that parallel the original C.A.T.</p> <p>3. Rorschach Inkblot Test</p> <p>Specification</p> <p>Rorschach Inkblot Test is one of the most widely used projective techniques for clinical assessment used by psychologists all over the world. It is most popular amongst various systems of interpretation both at teaching institutions in Clinical Psychology and at hospitals/clinics as well as with practicing mental health professionals. There are 10 cards in Rorschach ranging from black and white to multi coloured. subjects' perceptions of inkblots are recorded and then analyzed using psychological interpretation, complex scientifically derived algorithms, or both. Some psychologists use this test to examine a person's personality characteristics and emotional functioning. It has been employed to detect an underlying thought disorder, especially in cases where patients are reluctant to describe their thinking processes openly. The test takes its name from that of its creator, Swiss psychologist Hermann Rorschach. There are majorly three systems of scoring. 1. Beck's, 2. Klopfer's, 3. Exner's</p> <p>The complete set of Rorschach Ink blot test consists of cards, manual, record and analysis blank & location chart</p>	
PL - 47	Suction Machine	<p>Compact design</p> <ul style="list-style-type: none"> * Shock proof body * Noiseless technology * Motor 1/6 hp compressor. * Suction Capacity : 650 -700 mm /hg * Jar 2 * 1.5 glass capacity * Vacuum gauze 2.0 Inch * 0.760 mm/hg calibration * Power 200/240 Volts * 1440 rpm, 50 cycles 	

Terms and Conditions :

- 1) All rights are reserved by The Dean, C.P.R. General Hospital, Kolhapur to reject any or all quotations without assigning any reason.
- 2) Material quoted should be of best quality. Samples should be made available by the supplier for demonstration at free of cost to the user department or concern purchase committee. (If required).
- 3) Sealed quotation should reach to this office on or before Date :- 30/07/2020 within office time positively. Quotation received after above mentioned date will not be entertained. This office is not liable for any delay of Post Office or Courier Agency or at any other conditions. **Supplier has to enroll the quotation at inward section of this office and has to deposit the quotation in the box kept at Machinery Store Department, C.P.R. Hospital, Kolhapur. Quotation should be prepared on suppliers Letter Head only, quotation filled on downloaded print will not be accepted and will be treated as rejected.**
- 4) **Separate quotation should be submitted for each Machinery. Duly signed and stamped Quotation should be submitted on suppliers letter head only.**
- 5) If any of the above mentioned machinery, equipment is available of G.E.M., the purchase procedure for such machinery will be treated as cancelled.
- 6) The rates quoted should be inclusive of excise duty, GST, Transportation, Insurance, packing and forwarding and Fittings and Installation charges etc.. Rates should be within the market rate limits and should not be more than M.R.P. at any circumstances. At any stage of the quotation process even after completion of the process, if it is found that the rates mentioned are more than the M.R.P., the supplier is responsible for refund the difference with interest to this office.

7) If failed to supply the material as per required specification, then the order will be diverted to second lowest rate provider and in this case the difference between first lowest and second lowest will be born by first lowest provider.

8) Attach the self attested photo copies of PAN Card, GST Registration Certificate, Shop Registration Certificate (Shop Act License etc.)

9) Please superscript the envelope with "QUOTATION FOR Specification No. Name of Machinery "

Groz

Dean,

Chhatrapati Pramilaraje
General Hospital, Kolhapur.

1	उपि हल्लो मशिन	9	3840000		
2	सिग्नल लाइवर को.आर. सिस्टीम मशिन	9	9998800	9998800	Radiology
3	डाय.एन.डी. लेजर मशिन	9	2400000	2400000	Radiology
4	डाय.एन.डी. लेजर मशिन	9	1632600	1632600	Ophthalmology
5	अनस्थीया डेक स्टेशन (लो एअर)	9	9600000	3500000	Anesthesiology
6	सि.आन.मशिन	9	9542000	9542000	Surgery
7	सर्जिकल कॅन्टरी	9	9492000	9492000	Surgery
8	पेडियाट्रिक / मिथोनटल डेन्टीलेटर	9	9040000	2900000	Pediatrics
9	एलेक्ट्रोमिल फायबर ऑप्टिक डायग्नोस्टिक	9	9409824	9409824	TB & CHEST
10	कायनर ऑप्टिक डायग्नोस्टिक	9	9358932	9358932	Anesthesiology
11	पेडियाट्रिक डेन्टीलेटर इलेक्ट्रोमिल ऑपरेटड डेन्टील डेस्क	9	395680	633260	Dentistry
12	मल्टी से.मशिन	9	109200	696000	Surgery
13	पल्स ऑक्सिमिटर	98	82000	466000	Pediatrics
14	मोनिटिंग (मल्टीप्रीमरी & से.से.)	9	949200	608600	TB & CHEST
15	मल्टीप्रीमरीमशिन	9	94200	260600	Pediatrics
16	डाय.एन.डी. लेजर	9	299000	499960	Pediatrics
17	डाय.एन.डी. लेजर	9	299000	399960	Orthopedics
18	डाय.एन.डी. लेजर	9	368000	368000	Pediatrics
19	डाय.एन.डी. लेजर	9	368000	368000	ENT
20	डाय.एन.डी. लेजर	9	368000	368000	ENT
21	डाय.एन.डी. लेजर	9	368000	368000	TB & CHEST
22	डाय.एन.डी. लेजर	9	368000	368000	ENT